

STEAM ASSOCIATION OF MANITOBA

Steam Gauge Verification Record

Exhibit 12 (Rev A)

Gauge Identification

SAM SAM 22-05 Make Ashcroft Serial# 445921
 Gauge Max 300 Owner MAM GAAR SCOTT
 Engine/Boiler _____ MB BB# 6642 MAWP 120

Notes _____

Comparison

Date JUNE 23/22 Personnel Jim Down / ROBERT BEAMISH
 Lifts at (PSI) 12 Test max PSI 150 (MAWP × 1.25)

Rising Pressure				Falling Pressure			
Calibrated Gauge	Field Gauge	Calibrated Gauge	Field Gauge	Calibrated Gauge	Field Gauge	Calibrated Gauge	Field Gauge
10	12	160		10	6	160	
20	15	170		20	20	170	
30	26	180		30	32	180	
40	37	190		40	41	190	
50	48	200		50	50	200	
60	59	210		60	61	210	
70	67	220		70	70	220	
80	80	230		80	80	230	
90	90	240		90	91	240	
100	100	250		100	100	250	
110	110	260		110	112	260	
120	120	270		120	122	270	
130	129	280		130	130	280	
140	139	290		140	140	290	
150	148	300		150	148	300	

Revision Date
June 6, 2018