

STEAM ASSOCIATION OF MANITOBA

Steam Gauge Verification Record

Exhibit 12 (Rev A)

Gauge Identification

SA SAM 22-07 Make ASGV MC Serial# Y 1903
 Gauge Max 200 Owner IND CASE MAM
 Engine/Boiler _____ MB BB# 6650 MAWP 130
 Notes _____

Comparison

Date JUNE 23/22 Personnel JIM DOWN/ROBERT BEAMISK
 Lifts at (PSI) 5 Test max PSI 150 (MAWP × 1.25)

Rising Pressure				Falling Pressure			
Calibrated Gauge	Field Gauge	Calibrated Gauge	Field Gauge	Calibrated Gauge	Field Gauge	Calibrated Gauge	Field Gauge
10	10	160	158	10	12	160	158
20	21	170		20	22	170	
30	31	180		30	31	180	
40	41	190		40	41	190	
50	51	200		50	51	200	
60	61	210		60	61	210	
70	71	220		70	72	220	
80	81	230		80	82	230	
90	90	240		90	90	240	
100	100	250		100	100	250	
110	110	260		110	111	260	
120	120	270		120	120	270	
130	130	280		130	130	280	
140	139	290		140	140	290	
150	149	300		150	150	300	

Revision Date
June 6, 2018