

STEAM ASSOCIATION OF MANITOBA

Steam Gauge Verification Record

Exhibit 12 (Rev A)

Gauge Identification

SAM 22-10 Make WINTERS Serial# —
 Gauge Max 300 Owner DEVON ALLAN
 Engine/Boiler CASE 75 MB BB# — MAWP 131
 Notes —

Comparison

Date July 26 / 22 Personnel Jim Down / DEVON ALLAN
 Lifts at (PSI) 5 Test max PSI 170 (MAWP × 1.25)

Rising Pressure				Falling Pressure			
Calibrated Gauge	Field Gauge	Calibrated Gauge	Field Gauge	Calibrated Gauge	Field Gauge	Calibrated Gauge	Field Gauge
10	10	160	165	10	14	160	165
20	24	170	175	20	24	170	175
30	34	180		30	35	180	
40	44	190		40	45	190	
50	54	200		50	55	200	
60	64	210		60	65	210	
70	74	220		70	75	220	
80	84	230		80	85	230	
90	92	240		90	94	240	
100	104	250		100	104	250	
110	114	260		110	114	260	
120	124	270		120	124	270	
130	134	280		130	135	280	
140	144	290		140	145	290	
150	154	300		150	155	300	

Revision Date
June 6, 2018