

Wenig

Old School

Box 7

Rosalind Ab

TOB 3Y0

(780)608-4596

jawenig@xplornet.or

Instrument

Date: June 10/14

Pressure Gauge Calibration Report

Serial number: 533889

Manufacturer: Ashcroft

Dial size: 5"

Range: 300 psi

Connection: 3/8" nptm

Max error: 1.833 %

Testing Device: Budenberg 280L
Sn. 10928

Other: _____

Test medium: Air / Oil / N2

Ambient temperature: 20°C

Inspector: A. Wenig

Test pressure	Gauge reading
72.5 psi	78 psi
145 psi	149 psi
217.5 psi	217. psi
290 psi	290 psi