# STEAM ASSOCIATION OF MANITOBA

**Steam Gauge Verification Record** **Exhibit 12 (Rev A)**

**Gauge Identification**

SAM ID 24-01 Make ASHCROFT Serial# 445921

Gauge Max 300 lb. Owner MAM

Engine/Boiler Garr Scott MB BB# _____ MAWP 120 lbs.

Notes _____

**Comparison**

Date Feb 2, 2024 Personnel J. Down / K. Down

Lifts at (PSI) _____ Test max PSI 150 (MAWP × 1.25)

| Rising Pressure | | | | Falling Pressure | | | |
|---|---|---|---|---|---|---|---|
| FIELD ~~Calibrated~~ Gauge | MASTER ~~Field~~ Gauge | Calibrated Gauge | Field Gauge | FIELD ~~Calibrated~~ Gauge | MASTER ~~Field~~ Gauge | Calibrated Gauge | Field Gauge |
| 10 | 16 | 160 | | 10 | 12 | 160 | |
| 20 | 24 | 170 | | 20 | 20 | 170 | |
| 30 | 33 | 180 | | 30 | 28 | 180 | |
| 40 | 42 | 190 | | 40 | 39 | 190 | |
| 50 | 52 | 200 | | 50 | 49 | 200 | |
| 60 | 62 | 210 | | 60 | 58 | 210 | |
| 70 | 72 | 220 | | 70 | 69 | 220 | |
| 80 | 82 | 230 | | 80 | 78 | 230 | |
| 90 | 92 | 240 | | 90 | 88 | 240 | |
| 100 | 101 | 250 | | 100 | 98 | 250 | |
| 110 | 110 | 260 | | 110 | 108 | 260 | |
| 120 | 122 | 270 | | 120 | 117 | 270 | |
| 130 | 131 | 280 | | 130 | 129 | 280 | |
| 140 | 142 | 290 | | 140 | 139 | 290 | |
| 150 | 152 | 300 | | 150 | 152 | 300 | |