

STEAM ASSOCIATION OF MANITOBA

Steam Gauge Verification Record

Exhibit 12 (Rev D)

Gauge Identification

SAM ID 26-05 Make CASE/FEDRAL Serial# N/A.
 Gauge Max 300 Owner Down Family.
 Engine/Boiler CASE50 MB B# B6655 MAWP 150

Notes _____

Comparison

Date MAY 30/26 Personnel Jim Down / GORD Down
 Lifts at (PSI) 8 Test max PSI 190 (MAWP × 1.25)

Rising Pressure				Falling Pressure			
Field Gauge (Set)	Master Gauge (Read)	Field Gauge (Set)	Master Gauge (Read)	Field Gauge (Set)	Master Gauge (Read)	Field Gauge (Set)	Master Gauge (Read)
0	8	160	161	0	6	160	162
10	10	170	172	10	10	170	172
20	20	175	177	20	20	175	176
25	24	180	182	25	24	180	182
30	29	190	192	30	28	190	192
40	38	200		40	38	200	
50	50	210		50	48	210	
60	58	220		60	58	220	
70	68	225		70	68	225	
75	74	230		75	74	230	
80	79	240		80	79	240	
90	89	250		90	89	250	
100	99	260		100	100	260	
110	109	270		110	109	270	
120	120	275		120	120	275	
125	126	280		125	126	280	
130	130	290		130	130	290	
140	140	300		140	140	300	
150	150			150	151		

Revision Date
June 22, 2024