

STEAM ASSOCIATION OF MANITOBA

Steam Gauge Verification Record

Exhibit 12 (Rev D)

Gauge Identification

SAM ID 26-04 Make WINTERS Serial# AL6501
 Gauge Max 200 Owner MAM
 Engine/Boiler SOCASE MB B# B6644 MAWP 124

Notes _____

Comparison

Date MAY 30/26 Personnel Jim Down / Gordon Down
 Lifts at (PSI) 5 Test max PSI 150 (MAWP × 1.25)

Rising Pressure				Falling Pressure			
Field Gauge (Set)	Master Gauge (Read)	Field Gauge (Set)	Master Gauge (Read)	Field Gauge (Set)	Master Gauge (Read)	Field Gauge (Set)	Master Gauge (Read)
0	5	160	162	0		160	162
10	14	170		10	10	170	
20	21	175		20	18	175	
25		180		25		180	
30	28	190		30	29	190	
40	41	200		40	38	200	
50	49	210		50	48	210	
60	61	220		60	58	220	
70	69	225		70	70	225	
75		230		75		230	
80	81	240		80	78	240	
90	90	250		90	90	250	
100	101	260		100	99	260	
110	111	270		110	109	270	
120	121	275		120	119	275	
125		280		125		280	
130	128	290		130	131	290	
140	142	300		140	140	300	
150	148			150	150		

Revision Date
June 22, 2024