

STEAM ASSOCIATION OF MANITOBA

Steam Gauge Verification Record

Exhibit 12 (Rev D)

Gauge Identification

SAM ID 26-02 Make Jas Morrison ^{CSI} Serial# 024-352416
 Gauge Max 300 Owner M.A.M.
 Engine/Boiler LITTLE TINY white MB B# _____ MAWP _____
 Notes _____

Comparison

Date MAY 27/26 Personnel JIM DOWN / KAREN DOWN
 Lifts at (PSI) 5 Test max PSI 200 (MAWP × 1.25)

Rising Pressure				Falling Pressure			
Field Gauge (Set)	Master Gauge (Read)	Field Gauge (Set)	Master Gauge (Read)	Field Gauge (Set)	Master Gauge (Read)	Field Gauge (Set)	Master Gauge (Read)
0		160	161	0	5	160	159
10	12	170	170	10	10	170	168
20	21	175		20	19	175	
25		180	180	25		180	178
30	30	190	189	30	29	190	189
40	40	200	200	40	39	200	200
50	50	210		50	49	210	
60	61	220		60	59	220	
70	71	225		70	69	225	
75	80	230		75		230	
80	80	240		80	80	240	
90	89	250		90	88	250	
100	100	260		100	100	260	
110	110	270		110	109	270	
120	121	275		120	119	275	
125		280		125		280	
130	131	290		130	129	290	
140	141	300		140	139	300	
150	150			150	150		

Revision Date
June 22, 2024